

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

CB TEST CERTIFICATE

Product	Component power supply
Name and address of the applicant	BRIDGEPOWER CORP (GOSAEK-DONG) 16 OMOKCHEN-RO 132BEON-GIL GWONSEON-GU SUWON-SI GYEONGGI 441-813 KOREA
Name and address of the manufacturer	SL POWER ELECTRONICS CORP BLDG A 6050 KING DR VENTURA, CA 93003 USA
Name and address of the factory	BRIDGEPOWER CORP (GOSAEK-DONG) 16 OMOKCHEN-RO 132BEON-GIL GWONSEON-GU SUWON-SI GYEONGGI 441-813 KOREA <input checked="" type="checkbox"/> Additional Information on page 2
Note: <i>When more than one factory, please report on page 2</i>	
Ratings and principal characteristics	Rated Input: 100-240Vac, 50-60Hz, 1.2A (1.2A-0.5A) For only 5V output, Rated Input : 100-240Vac, 400Hz, 1.2A (1.2A-0.5A), See Page 2
Trademark (if any)	SL POWER ELECTRONICS
Type of Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	GB30(1)(2)(3)(4), GB40(1)(2)(3)(4), See Page 2
Additional information (if necessary may also be reported on page 2)	Class I <input checked="" type="checkbox"/> Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012
As shown in the Test Report Ref. No. which forms part of this Certificate	E302267-D1037-1/A0/C0-CB issued on 2018-07-06

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2018-07-23

Signature:

Jan-Erik Storgaard



Ref. Certif. No.

DK-75013-UL

Model Details:

GB30(1)(2)(3)(4), GB40(1)(2)(3)(4)

(Where (1), (3), (4) may alphanumeric, "For marketing purpose and no impact safety related critical components and constructions", where (2) may any number 05 through 48)

(1) Model Configuration : A to Z (Standard)

(2) Output Voltage : 05, 07, 09, 12, 15, 18, 24, 48

(3) Standard Input Connector Options Can be K or P for input type. Photographs for each plug-type configuration K (Class I = Connector type) P : (Class I = Pin type)

(4) Model Configuration : Number : 00 thru 99

Factories:

WENDENG JEIL ELECTRONICS CO LTD

XIAMEN ROAD NO.2 WENDENG ECONOMIC DEVELOPMENT ZONE WEIHAI CITY SHANDONG PROVINCE CHINA

BRIDGEPOWER VINA CO LTD

LOT B9 THUY VAN INDUSTRIAL ZONE VIET TRI CITY PHU THO PROVINCE VIETNAM

Additional Information:

Risk Management was not included in this investigation.

Additionally evaluated to EN 60601-1:2006/ A1:2013/ A12:2014. National Differences specified in the CB Test Report.

Additional information (if necessary)



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2018-07-23

Signature:

Jan-Erik Storgaard